



INSTRUCTION MANUAL

MICROAIRE®
PULSELAVAGE

INDICATIONS FOR USE

The MicroAire® Pulse Lavage is a single-use disposable pulse lavage system that provides controlled pulsatile irrigation and suction. MicroAire® Pulse Lavage is used in general osteotomies, for wound debridement, and for cleansing of surgical sites.

A variety of single-use disposable tips and splash shields are offered for MicroAire® Pulse Lavage.

APPLICABLE PART NUMBERS

KITS

REF Number	DESCRIPTION	UOM
201-02-00	MicroAire® Pulse Lavage Component Kit (includes disposable instrument, tubing, battery pack)	6/box
201-03-00	Knee Kit (includes MicroAire® Pulse Lavage Component Kit and shower spray tip with splash shield)	6/box
201-04-00	Hip Kit (includes MicroAire® Pulse Lavage Component Kit, femoral spray tip and shower spray tip with splash shield)	6/box
201-05-00	Fan Spray Kit (includes MicroAire® Pulse Lavage Component Kit and fan spray tip with splash shield)	6/box

TIPS

REF Number	DESCRIPTION	UOM
201-00-37	Shower spray tip with splash shield	6/box
201-00-44	Fan spray tip with splash shield	6/box
201-00-46	Femoral spray tip	6/box
201-00-50	Shower Spray tip	6/box
201-00-51	Fan spray tip	6/box
201-00-52	Femoral brush tip	6/box

SPLASH SHIELDS

REF Number	DESCRIPTION	UOM
201-00-80	Splash shield, 6 inch (15.2 cm)	6/box

SYMBOL DEFINITIONS



Consult Instruction for Use



Sterile Method of Sterilization



European Conformity Mark



Date of Manufacture, YYYY-MM.

REF

Product Catalog Number



Use By Date, YYYY-MM



Product Lot Number



Temperature Limitations



Humidity Limitations



Authorized European Representative



Tip Locked



Tip Unlocked



Normal Operation



High Pulsatile Operation



Do Not Immerse



Type BF Applied Part



Single Patient Use Only -Do not resterilize

**SHORT
TIME
OPERATION**

Short time operation, maximum 8 minutes



Contents sterile only if package and seal are undamaged and product has not exceeded its Use By Date



cTUVus Certification

INTRODUCTION

Throughout the manual, the following terms are used to help identify tips and precautions that will help avoid accidental injury to patients or personnel, or prevent damage to the system.

WARNING: Used to indicate that the safety of the patient and hospital personnel could be involved.

⚠ GENERAL WARNINGS

WARNING Visually inspect for damage to the package and product prior to use. Do not use the product if package integrity has been compromised.

NOTE 1 If concerns are noted that may compromise the function of the device, do not use the product and please contact your representative.

NOTE 2 No absolute Contraindications exist.

CAUTION

- Federal Law restricts this device to sale by or on the order of a physician (or properly licensed practitioner).

CAUTION: Used to point out special procedures or precautions that must be followed to avoid damaging the system / instrument.

NOTE: Used to point out or suggest the easiest means of carrying out techniques.

- Read Instructions For Use prior to the operation of MicroAire® Pulse Lavage.
- The MicroAire® Pulse Lavage needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in the manual.
- Portable and mobile RF communications equipment can affect Medical Electrical Equipment.

INSTRUCTIONS FOR USE

1. Remove MicroAire® Pulse Lavage from packaging tray by peeling the Tyvek blister (Figure A).




Figure A

2. Connect MicroAire® Pulse Lavage irrigation tubing to an irrigation bag by inserting the bag spike into the bag's connection valve. Put irrigation bag no more than 0.8 meters or 2.6 feet above the handpiece (Figure B).




Figure B

3. Connect MicroAire® Pulse Lavage suction tubing to a suction canister by placing the blue suction connector onto the barbed canister fitting (Figure B).
4. Pull up the tip lock to the unlocked position to prepare it for tip insertion (Figure C).




Figure C

5. Insert the desired tip into MicroAire® Pulse Lavage (Figure D).




Figure D

- Push down the tip lock to the locked position to secure the desired tip in place (Figure E).




Figure E

- Prime MicroAire® Pulse Lavage by depressing the trigger to either the normal (▶) or high pulsatile position (▶▶) for a few seconds while aiming the tip towards the surgical site (Figure F).




Figure F

- Irrigate the surgical site as needed. To limit irrigation flow, or to prevent dripping, tighten the tubing clamp that is attached to the irrigation tubing.
- Release or depress the pinch clamp on the suction tubing to control suction.
- To change tip, turn off MicroAire® Pulse Lavage, pull up and release tip lock. Pull out existing tip and replace with desired tip.

WARNINGS

- Equipment is not suitable for use in the presence of a flammable anaesthetic mixture with air or with oxygen or nitrous oxide.

- Universal precautions for handling contaminated materials should be observed at all times.
- DO NOT resterilize the instrument. MicroAire® Pulse Lavage is a single-use disposable device.
- DO NOT cut the battery case from the instrument or remove batteries from the battery case. Doing so may produce heat and/or electric shock, which may lead to personal injury and/or fire. The case contains mercury-free alkaline batteries which do not require specific disposal methods in most areas. We do not recommend that the batteries be removed from their pack. Batteries can cause serious burns to both the skin and eyes.



NOTE

- Prior to use, make sure that the tip lock is fully engaged and the tip is locked to the instrument. Failure to do so may cause the tip to fall off during use.
- Suction pressure may be reduced by clogs or debris in the suction path. For best performance, keep suction path clear of obstructions.
- Weakened performance after prolonged use may be the result of battery drainage.
- MicroAire® Pulse Lavage is intended to be used only by trained medical professionals who are familiar with its use and application.
- Tips may have limited contact with the surgical sites during operation.
- Positioning irrigation bag below the handpiece may weaken the performance or make the instrument malfunction.


ENVIRONMENTAL PARAMETERS

CAUTION Sterile, packaged devices should be stored in a designated, limited access area that is well ventilated and provides protection from dust, moisture, insects, vermin, and temperature/humidity extremes.

OPERATING CONDITIONS



SHIPPING AND STORAGE CONDITIONS



The materials and components used in the construction of this device were selected to ensure that the device could be shipped by any standard commercial method without special handling conditions. ORDINARY (IPXO) PROTECTION. INTERNALLY POWERED DEVICE.

LIMITED WARRANTY MicroAire® Surgical instruments warrants to the original purchaser/consumer that this product will be free from manufacturing defects in material and workmanship for a period of 90 days from the original purchase date. The warranty is limited and only guarantees the replacement of the product without charge to the original purchaser/consumer, provided

that the faulty handpiece is returned. This warranty is non-transferrable.

In the event of abuse, misuse, or use for anything other than the manner described in the written instructions furnished by the manufacturer, this limited warranty shall be null and void.

RECOMMENDED SEPARATION DISTANCES BETWEEN PORTABLE AND MOBILE RF COMMUNICATIONS EQUIPMENT AND THE MicroAire® PULSE LAVAGE.

The MicroAire® Pulse Lavage is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the MicroAire® Pulse Lavage can help prevent electromagnetic interference by maintaining a minimum distance

between portable and mobile RF equipment (transmitters) and the MicroAire® Pulse Lavage as recommended below, according to the maximum output power of the communication equipment.

Rated maximum output power of transmitter. WATTS	Separation distance according to the frequency of transmitter. METERS		
	150 kHz to 80 MHz $d=1.2 \sqrt{P}$	80 MHz to 800 MHz $d=1.2 \sqrt{P}$	800 MHz to 2.5 GHz $d=2.3 \sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

IEC 60601-1-2 Table 206

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

GUIDANCE AND MANUFACTURER'S DECLARATION

VIBRATION AND NOISE EMISSION

Vibration Exposure	a_{hv} (m/s ²)	1.22
	Uncertainty K (m/s ²)	-
Noise Emission Value	$L_{PA}(dB(A))$	75

ELECTROMAGNETIC EMISSIONS

The MicroAire® Pulse Lavage is intended for use in the electromagnetic environment specified below. The customer or the user of the

MicroAire® Pulse Lavage should assure that it is used in such an environment.

EMISSIONS TEST	COMPLIANCE	ELECTROMAGNETIC ENVIRONMENT - GUIDANCE
Harmonic emissions IEC 61000-3-2	Not applicable	The MicroAire® Pulse Lavage is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Voltage fluctuations / flicker emissions IEC 61000-3-3	Not applicable	
RF emissions CISPR 14-1	Complies	

IEC 61000-3-2 Table 201

WARNING: This Pulse Lavage should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, the equipment should be observed to verify normal operation in the configuration in which it will be used.

WARNING: The use of accessories, transducers, and cable other than those specified by the manufacturer, may result in increased emissions or decreased immunity of the MicroAire® Pulse Lavage.


ELECTROMAGNETIC IMMUNITY

The MicroAire® Pulse Lavage is intended for use in the electromagnetic environment specified below. The customer or the user of

the MicroAire® Pulse Lavage should assure that it is used in such an environment.

IMMUNITY TEST	IEC 60601 TEST LEVEL	COMPLIANCE LEVEL	ELECTROMAGNETIC ENVIRONMENT – GUIDANCE
Electrostatic discharge (ESD). IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/ burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	Not applicable. The device is battery powered. ±1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV for common mode	Not applicable. The device is battery powered.	Mains power quality should be that of a typical commercial or hospital environment.

IMMUNITY TEST	IEC 60601 TEST LEVEL	COMPLIANCE LEVEL	ELECTROMAGNETIC ENVIRONMENT – GUIDANCE
Voltage dips, short interruptions and voltage variations on power supply input lines. IEC 61000-4-11	<5% U (95% dip in U) for 0, 5 cycle 40% U (60% dip in U) for 5 cycles 70% U (30% dip in U) for 25 cycles <5% U (>95% dip in U) for 5 seconds	Not applicable.	Mains power quality should be that of a typical commercial or hospital environment. If the user of the MicroAire® Pulse Lavage requires continued operation during power mains interruptions, it is recommended that the MicroAire® Pulse Lavage be powered from an uninterruptible power supply or battery. The device is battery powered.
Power frequency (50/60 Hz) magnetic field. IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms	<p>Portable and mobile RF communications equipment should be used no closer to any part of the MicroAire® Pulse Lavage, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended Separation distance- $d=1.2 \sqrt{P}$ $d=1.2 \sqrt{P}$ - 80 MHz to 800 MHz $d=2.3 \sqrt{P}$ - 800 MHz to 2.5 GHz</p> <p>Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strength from fixed RF transmitters, as determined by an electromagnetic site survey, should be less than the compliance level in each frequency range.</p> <p>Interference may occur in the vicinity of equipment marked with following symbol:</p> 
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m	

IEC 60601-1-2 Table 204

NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Field strengths from fixed transmitters, such as base stations for radio (cellular / cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the

electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the MicroAire® Pulse Lavage is used exceeds the applicable RF compliance level above, the MicroAire® Pulse Lavage should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the MicroAire® Pulse Lavage. Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.



Manufactured for:

MicroAire Surgical Instruments LLC

3590 Grand Forks Boulevard

Charlottesville, VA 22911 USA

800-722-0822

www.microaire.com



MediMark Europe

11, rue Emile Zola - BP 2332

F-38033 Grenoble Cedex 2

France



Manufactured for MicroAire by Apex Tools & Orthopedics Co., No.25, Yonghua Road, Yonghe, GETDD, Guangzhou, PRC, 511356.

Made in China.